

Radium Delayed Coincidence Counter (Scientific Computer Instruments); 8-Channel Calibrations:

System	Efficiency, Ra-223 (cpm/dpm)	Efficiency, Ra-224 (cpm/dpm)	Efficiency, Total (cpm/dpm)
1	0.357	0.492	1.128
2	0.350	0.500	1.151
3	0.360	0.496	1.142
4	0.334	0.461	1.101
5	0.303	0.421	0.978
6	0.347	0.482	1.119
7	0.364	0.505	1.174
8	0.359	0.499	1.158

Minimum detectable activities for this data set average ( $\pm$  standard deviation):  $5.4 \pm 12.9$  dpm/100L for Ra-223 and  $4.0 \pm 18.4$  dpm/100L for Ra-224 and Ra-228 from water column samples (CTD, Core-top waters, and Alvin Niskins)

Minimum detectable activities for this data set average ( $\pm$  standard deviation):  $61 \pm 45$  dpm/100L for Ra-223 and  $39 \pm 29$  dpm/100L for Ra-224 and Ra-228 from Porewater samples.

Radon Emanation Line; 4-port

Calibration curves:

$$\text{Port 1: dpm} = 0.0009\text{cpm}^2 + 0.4389\text{cpm} + 0.8736$$

$$\text{Port 2: dpm} = 0.0009\text{cpm}^2 + 0.4169\text{cpm} + 1.0583$$

$$\text{Port 3: dpm} = 0.0004\text{cpm}^2 + 0.5437\text{cpm} + 0.9103$$

$$\text{Port 4: dpm} = 0.0004\text{cpm}^2 + 0.4676\text{cpm} + 0.9789$$

Minimum detectable activities for this data set average ( $\pm$  standard deviation):  $24.7 \pm 49.2$  dpm/100L for Ra-226.